

99009040261000

Report the operation or significant modification of technical X-ray equipment

Heruntergeladen am 16.07.2025

<https://fimportal.de/xzufi-services/281632405/L100012>

Modul	Sachverhalt
Leistungsschlüssel	99009040261000
Leistungsbezeichnung I	Report the operation or significant modification of technical X-ray equipment
Leistungsbezeichnung II	Report the operation or significant modification of technical X-ray equipment
Typisierung	2 - Bundesauftragsverwaltung: Regelung
Quellredaktion	Schleswig-Holstein
Freigabestatus Katalog	fachlich freigegeben (gold)
Freigabestatus Bibliothek	fachlich freigegeben (gold)
Begriffe im Kontext	
Leistungstyp	Leistungsobjekt mit Verrichtung
Leistungsgruppierung	Atomare Angelegenheiten (009)
Verrichtungskennung	Entgegennahme (261)
SDG-Informationsbereich	Vorschriften für und Anforderungen an Erzeugnisse

Modul	Sachverhalt
Lagen Portalverbund	Anmeldepflichten (2010100), Prüfung und Nachweise für Sachkunde und Sicherheit (2120300), Anlagenbetrieb und -prüfung (2120100)
Einheitlicher Ansprechpartner	Nein
Fachlich freigegeben am	31.01.2024
Fachlich freigegeben durch	Federal Ministry for the Environment, Nature Conservation, Nuclear Safety and Consumer Protection (BMUV)
Handlungsgrundlage	https://www.gesetze-im-internet.de/strlrschg/_19.html
Teaser	Would you like to operate a technical X-ray device that does not require a permit or make significant changes to it? Then you must notify the competent authority for radiation protection.
Volltext	Technical X-ray equipment that does not require a permit may include X-ray equipment that is to be used as basic, high or full protection equipment. You must notify the competent authority in writing of the planned commissioning or significant modification. Significant changes to an X-ray device can be, for example • Change of room • Structural changes to the room • Change of the image receiver
Erforderliche Unterlagen	• Proof of expertise in radiation protection in accordance with § 74 Radiation Protection Act in conjunction with § 47 Radiation Protection Ordinance. • § 47 Radiation Protection Ordinance with proof of updating • Certificate and test report from an expert on the radiation protection test • Proof of knowledge of radiation protection and of updates to this knowledge for personnel working at the X-ray facilities If necessary, you must also submit

Modul	Sachverhalt
	• Type approval certificate with confirmation of routine testing • CE certificate of conformity
Voraussetzungen	You wish to operate or significantly modify a technical X-ray device that does not require a permit.
Kosten	
Verfahrensablauf	
Bearbeitungsdauer	2 - 4 Woche(n) If additional documents have to be requested and submitted, the processing time will be extended.
Frist	4 Woche(n) You must submit a notification with the necessary documents to the competent authority at least 4 weeks before the planned commissioning or significant change to the operation of the X-ray equipment.
weiterführende Informationen	
Hinweise	There are no indications or special features.
Rechtsbehelf	Action before the administrative court
Kurztext	• Notification of the operation of an X-ray device or a significant change to the operation Receipt • Companies with technical X-ray equipment must notify the competent authority if they wish to put the X-ray equipment into operation for the first time wish to make significant changes to the operation of the X-ray equipment, e.g: change of room structural changes to the room change of the image receiver • Notification must be made even if the X-ray equipment is not subject to approval • The following documents are required for notification: Proof of expertise in radiation protection with proof of updating Certificate and test report from the expert on the radiation protection test Proof of knowledge of radiation protection and of updates to this knowledge for the personnel working on the X-ray equipment Type approval certificate with confirmation of routine testing, if applicable CE certificate of conformity, if applicable

Modul	Sachverhalt
	• Notification must be made in writing • Responsible: competent authority for radiation protection
Ansprechpunkt	
Zuständige Stelle	
Formulare	
Ursprungsportal	Report the operation or significant modification of technical X-ray equipment, Betrieb oder die wesentliche Änderung einer technischen Röntgeneinrichtung anzeigen